



California Medical Device Recall Information



Recall Name

GE Healthcare Recalls Giraffe and Panda T-Piece/Bag and Mask Resuscitation Systems Due to a Potential Manufacturing and/or Labeling Error

Recall Date	Product Description	Recalling Firm	Recall Reason
10/15/12	Giraffe and Panda T-Piece Resuscitation System and Giraffe and Panda Bag and Mask Resuscitation System	GE Healthcare, LLC. Waukesha, WI	<i>Potentially incorrect oxygen and air wall inlet fittings and/or labels, both located on the back panel of the resuscitation systems;</i> <i>The components may have been mistakenly reversed during assembly.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Giraffe and Panda T-Piece Resuscitation System; <i>and</i> Giraffe and Panda Bag and Mask Resuscitation System Suspected Models Recalled: <ul style="list-style-type: none">• M1139417• M1139612• M1223628• M1192226-072951• M1192226-084200• M1192226-095181• M1192226-104599	CA , nationwide	Manufactured from April 1, 2008 through September 30, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm339868.htm>